

## EUROPEAN PATENT OFFICE

## Patent Abstracts of Japan

PUBLICATION NUMBER : 2001046364  
 PUBLICATION DATE : 20-02-01


APPLICATION DATE : 09-08-99  
 APPLICATION NUMBER : 11225681

APPLICANT : HITACHI MEDICAL CORP;

INVENTOR : YOSHIOKA TOMOTSUNE;

INT.CL. : A61B 6/03 G01T 1/20

TITLE : X-RAY DETECTOR AND X-RAY CT  
 DEVICE USING THE SAME



**ABSTRACT :** PROBLEM TO BE SOLVED: To provide an X-ray detector and an X-ray CT device using it capable of minimizing the error of the position deviation between a slice thickness direction collimator plate end part and the slice thickness direction of an X-ray detection element block separation layer, improving X-ray utilization efficiency and homogenizing detection characteristics among respective X-ray detection elements.

**SOLUTION:** A collimator structure body composed of a channel direction collimator 4 and a slice thickness direction collimator 5 and plural X-ray detection element columns composed of a scintillator 71 for making X-rays transmitted through a testee body 3 be incident and emitting light and a photoelectric conversion element for generating output signals corresponding to an X-ray quantity detected by the scintillator are arrayed for plural columns respectively in a channel direction and the slice thickness direction on a circular arc centering an X-ray tube focus. In this case, the width of a separation band for separating the X-ray detection element columns of the slice thickness direction of this X-ray detector 8 is made smaller than the width of the collimator plate of the slice thickness direction collimator.

COPYRIGHT: (C)2001,JPO

BEST AVAILABLE COPY